


CERTIFICATE OF ANALYSIS

General information

| | |
|---------------------|--|
| Name | HRD-P10-N(normal) |
| Cat. No. | CBP90023-10 |
| Lot No. | CBFA25051513X |
| Format | Genomic DNA |
| Buffer | Tris-EDTA |
| Storage Conditions | 2~8℃ |
| Date of Manufacture | 2025/05/15 |
| Expiry | 36 months from the date of manufacture |

Quality Control

| Test Items | Standard | Results | Passed |
|--------------------------------------|---------------------|---|--------|
| Quantity | 0.50 µg/vial*1 vial | 0.50 µg/vial*1 vial | √ |
| Concentration | 40.00 ng/uL±10% | 43.40 ng/uL by Qubit 3.0 detection | √ |
| OD ₂₆₀ /OD ₂₈₀ | 1.7~2.1 | 1.78 by Nanodrop detection | √ |
| DNA electrophoresis | No Degradation |  <p>M: Marker 1: HRD-P10-N(normal)</p> | √ |

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 彭亚楠 Date:2025/05/23

刘明 Date:2025/05/23