


CERTIFICATE OF ANALYSIS

General information

| | |
|---------------------|---|
| Name | HRD-P15-T (Tumor) |
| Catalog No. | CBP90023-15 |
| Lot No. | CBFA23061303 |
| HRD Score | Critical Value 45= 16(LOH)+20(TAI)+9(LST) with Panel+SNP method; Critical Value 43=9(LOH)+15(TAI)+19(LST) with WGS method; |
| Format | Genomic DNA |
| Buffer | Tris-EDTA |
| Storage Conditions | 4°C |
| Date of Manufacture | 2023/06/13 |
| Expiry | 36 months from the date of manufacture |

Quality Control

| Test Items | Standard | Results | Passed |
|--------------------------------------|---------------------|---|--------|
| Quantity | 1.00 µg/vial*1 vial | 1.00 µg/vial*1 vial | √ |
| Concentration | 40.00 ng/uL±10% | 43.80 ng/uL by Qubit 3.0 detection | √ |
| OD ₂₆₀ /OD ₂₈₀ | 1.7~2.1 | 1.83 by Nanodrop detection | √ |
| DNA electrophoresis | No Degradation |  <p>M: Marker 1: HRD-P15-T</p> | √ |

Caution:

For research use only.

If you have any questions about the Certificate of Analysis, please contact us.

Certified by: 陈毅 Date:2024/07/23

孙一扬 Date:2024/07/23